

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022023\
 Data File : VX034188.D
 Acq On : 20 Feb 2023 15:24
 Operator : JC/MD
 Sample : 01576-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-BAGFILTER-20220215

Quant Time: Feb 21 01:08:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	351475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	584685	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	521593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	241050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	207418	43.819	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	87.640%	
35) Dibromofluoromethane	5.385	113	183839	48.292	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.580%	
50) Toluene-d8	8.647	98	700609	51.259	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	102.520%	
62) 4-Bromofluorobenzene	11.079	95	250086	47.835	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	95.680%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.941	59	30676	28.975	ug/l	98
16) Acetone	2.374	43	66640	27.291	ug/l #	89
20) Methylene Chloride	2.788	84	24113	5.481	ug/l	97
37) Ethyl Acetate	4.715	43	73611	11.699	ug/l	99
43) Isopropyl Acetate	6.336	43	303338	27.320	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022023\
Data File : VX034188.D
Acq On : 20 Feb 2023 15:24
Operator : JC/MD
Sample : 01576-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-BAGFILTER-20220215

Quant Time: Feb 21 01:08:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

