

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
 Data File : VX040346.D
 Acq On : 20 Feb 2024 11:10
 Operator : JC/MD
 Sample : VX0220WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0220WBL01

Quant Time: Feb 21 01:09:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81232	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	139838	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135934	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61159	50.608	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.220%
35) Dibromofluoromethane	5.379	113	46797	47.921	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.840%
50) Toluene-d8	8.647	98	174184	47.683	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.360%
62) 4-Bromofluorobenzene	11.079	95	66522	44.651	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.300%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
Data File : VX040346.D
Acq On : 20 Feb 2024 11:10
Operator : JC/MD
Sample : VX0220WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0220WBL01

Quant Time: Feb 21 01:09:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

