

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
 Data File : VX040350.D
 Acq On : 20 Feb 2024 12:45
 Operator : JC/MD
 Sample : P1534-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-TB-20240216

Quant Time: Feb 21 01:12:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	76700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	138607	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132889	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60248	52.800	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.600%
35) Dibromofluoromethane	5.385	113	45184	46.680	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.360%
50) Toluene-d8	8.647	98	170221	47.012	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.020%
62) 4-Bromofluorobenzene	11.079	95	66658	45.139	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.280%
Target Compounds						
16) Acetone	2.374	43	1054	2.249	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
Data File : VX040350.D
Acq On : 20 Feb 2024 12:45
Operator : JC/MD
Sample : P1534-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-TB-20240216

Quant Time: Feb 21 01:12:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

