

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
 Data File : VX040351.D
 Acq On : 20 Feb 2024 13:07
 Operator : JC/MD
 Sample : P1507-22DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-20240215DL

Quant Time: Feb 21 01:12:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	78940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136948	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65774	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59932	51.032	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.060%
35) Dibromofluoromethane	5.385	113	45691	47.775	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.560%
50) Toluene-d8	8.647	98	169786	47.459	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.920%
62) 4-Bromofluorobenzene	11.079	95	65742	45.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.120%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	1094	1.307	ug/l	93
12) 1,1-Dichloroethene	2.319	96	345	0.428	ug/l #	66
44) Trichloroethene	7.123	130	67146	64.974	ug/l	99
64) Tetrachloroethene	9.268	164	423	0.454	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
Data File : VX040351.D
Acq On : 20 Feb 2024 13:07
Operator : JC/MD
Sample : P1507-22DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-20240215DL

Quant Time: Feb 21 01:12:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

