

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
 Data File : VX040354.D
 Acq On : 20 Feb 2024 14:16
 Operator : JC/MD
 Sample : P1414-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB020724

Quant Time: Feb 21 01:13:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	77362	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	138093	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60967	52.973	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.940%
35) Dibromofluoromethane	5.385	113	45194	46.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.720%
50) Toluene-d8	8.647	98	168576	46.730	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.460%
62) 4-Bromofluorobenzene	11.079	95	63146	42.920	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.840%
Target Compounds						
71) Bromoform	10.799	173	569	2.686	ug/l	Qvalue # 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
Data File : VX040354.D
Acq On : 20 Feb 2024 14:16
Operator : JC/MD
Sample : P1414-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB020724

Quant Time: Feb 21 01:13:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

