

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
 Data File : VX040356.D
 Acq On : 20 Feb 2024 15:01
 Operator : JC/MD
 Sample : P1521-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20240215

Quant Time: Feb 21 01:14:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	132626	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127231	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63779	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58049	52.689	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.380%
35) Dibromofluoromethane	5.385	113	43391	46.849	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.700%
50) Toluene-d8	8.647	98	162232	46.826	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.660%
62) 4-Bromofluorobenzene	11.079	95	61795	43.733	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.460%
Target Compounds						
16) Acetone	2.373	43	1074	2.374	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
Data File : VX040356.D
Acq On : 20 Feb 2024 15:01
Operator : JC/MD
Sample : P1521-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20240215

Quant Time: Feb 21 01:14:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

