

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
 Data File : VX040362.D
 Acq On : 20 Feb 2024 17:19
 Operator : JC/MD
 Sample : P1520-04 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20240215

Quant Time: Feb 21 01:16:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	70582	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	125051	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64195	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56294	53.611	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	107.220%	
35) Dibromofluoromethane	5.391	113	41198	47.176	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.360%	
50) Toluene-d8	8.647	98	154425	47.272	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	94.540%	
62) 4-Bromofluorobenzene	11.079	95	60498	45.409	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	90.820%	
Target Compounds						
						Qvalue
44) Trichloroethene	7.129	130	31828	33.729	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
Data File : VX040362.D
Acq On : 20 Feb 2024 17:19
Operator : JC/MD
Sample : P1520-04 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20240215

Quant Time: Feb 21 01:16:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

