

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
 Data File : VX040363.D
 Acq On : 20 Feb 2024 17:42
 Operator : JC/MD
 Sample : P1520-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP04-20240215

Quant Time: Feb 21 01:16:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70718	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	126905	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56886	54.070	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.140%
35) Dibromofluoromethane	5.385	113	42531	47.991	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.980%
50) Toluene-d8	8.647	98	157884	47.625	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.260%
62) 4-Bromofluorobenzene	11.079	95	61175	45.246	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.500%
Target Compounds						
44) Trichloroethene	7.135	130	764	0.798	ug/l	Qvalue 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022024\
Data File : VX040363.D
Acq On : 20 Feb 2024 17:42
Operator : JC/MD
Sample : P1520-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP04-20240215

Quant Time: Feb 21 01:16:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

