

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022119\
 Data File : VX007623.D
 Acq On : 21 Feb 2019 11:49
 Operator : JC/SP
 Sample : PB117256TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117256TB

Quant Time: Feb 22 04:59:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	466528	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	737759	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	716500	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	334315	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	262002	52.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.82%	
35) Dibromofluoromethane	5.51	113	245109	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
50) Toluene-d8	8.71	98	899088	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
62) 4-Bromofluorobenzene	11.13	95	311530	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
Target Compounds						
16) Acetone	2.45	43	28281	12.468	ug/l	94
20) Methylene Chloride	2.86	84	14581	2.776	ug/l	94
25) 2-Butanone	4.70	43	6250	1.936	ug/l #	80
43) Isopropyl Acetate	6.47	43	26498	2.566	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022119\
Data File : VX007623.D
Acq On : 21 Feb 2019 11:49
Operator : JC/SP
Sample : PB117256TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117256TB

Quant Time: Feb 22 04:59:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

