

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022119\
 Data File : VX007628.D
 Acq On : 21 Feb 2019 14:02
 Operator : JC/SP
 Sample : PB117250ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117250ZHE#02

Quant Time: Feb 22 05:47:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	461720	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	746108	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	721130	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	341102	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	264154	53.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.80%	
35) Dibromofluoromethane	5.50	113	245333	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
50) Toluene-d8	8.71	98	912803	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
62) 4-Bromofluorobenzene	11.13	95	321982	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
Target Compounds						
16) Acetone	2.45	43	31952	14.233	ug/l	97
20) Methylene Chloride	2.86	84	7455	1.242	ug/l #	93
43) Isopropyl Acetate	6.46	43	24566	2.352	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022119\
Data File : VX007628.D
Acq On : 21 Feb 2019 14:02
Operator : JC/SP
Sample : PB117250ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117250ZHE#02

Quant Time: Feb 22 05:47:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

