

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022119\
 Data File : VX007629.D
 Acq On : 21 Feb 2019 14:29
 Operator : JC/SP
 Sample : PB117250ZHE#03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117250ZHE#03

Quant Time: Feb 22 05:50:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	458137	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	739521	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	710843	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	331950	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	257748	53.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.00%	
35) Dibromofluoromethane	5.51	113	246746	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
50) Toluene-d8	8.71	98	899358	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
62) 4-Bromofluorobenzene	11.13	95	313290	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
Target Compounds						
16) Acetone	2.45	43	31017	13.925	ug/l	90
18) Methyl Acetate	2.78	43	10426	2.368	ug/l	99
20) Methylene Chloride	2.86	84	6468	1.036	ug/l #	80
43) Isopropyl Acetate	6.46	43	24470	2.364	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022119\
Data File : VX007629.D
Acq On : 21 Feb 2019 14:29
Operator : JC/SP
Sample : PB117250ZHE#03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117250ZHE#03

Quant Time: Feb 22 05:50:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

