

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022119\
 Data File : VX007633.D
 Acq On : 21 Feb 2019 16:16
 Operator : JC/SP
 Sample : PB117250ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117250ZHE#07

Quant Time: Feb 22 05:59:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	440218	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	710087	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	698447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	322608	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	249833	53.47	ug/l	0.00
Spiked Amount			Recovery	=	106.94%	
35) Dibromofluoromethane	5.50	113	231819	50.17	ug/l	0.00
Spiked Amount			Recovery	=	100.34%	
50) Toluene-d8	8.71	98	871716	50.85	ug/l	0.00
Spiked Amount			Recovery	=	101.70%	
62) 4-Bromofluorobenzene	11.13	95	305488	48.81	ug/l	0.00
Spiked Amount			Recovery	=	97.62%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	28833	13.471	ug/l	100
18) Methyl Acetate	2.78	43	11740	2.673	ug/l #	82
20) Methylene Chloride	2.86	84	12028	2.377	ug/l	96
25) 2-Butanone	4.71	43	5854	1.922	ug/l #	82
43) Isopropyl Acetate	6.46	43	25646	2.580	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022119\
Data File : VX007633.D
Acq On : 21 Feb 2019 16:16
Operator : JC/SP
Sample : PB117250ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117250ZHE#07

Quant Time: Feb 22 05:59:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

