

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022119\
 Data File : VX007639.D
 Acq On : 21 Feb 2019 18:56
 Operator : JC/SP
 Sample : PB117250ZHE#13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117250ZHE#13

Quant Time: Feb 22 06:17:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	434865	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	713020	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	705453	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	331409	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	248446	53.83	ug/l	0.00
Spiked Amount			Recovery	=	107.66%	
35) Dibromofluoromethane	5.50	113	234092	50.46	ug/l	0.00
Spiked Amount			Recovery	=	100.92%	
50) Toluene-d8	8.71	98	879619	51.10	ug/l	0.00
Spiked Amount			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	11.13	95	314770	50.09	ug/l	0.00
Spiked Amount			Recovery	=	100.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	25745	12.176	ug/l	97
18) Methyl Acetate	2.78	43	13591	3.031	ug/l	98
20) Methylene Chloride	2.86	84	12937	2.623	ug/l	92
25) 2-Butanone	4.70	43	5136	1.707	ug/l	94
43) Isopropyl Acetate	6.46	43	25159	2.520	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022119\
Data File : VX007639.D
Acq On : 21 Feb 2019 18:56
Operator : JC/SP
Sample : PB117250ZHE#13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117250ZHE#13

Quant Time: Feb 22 06:17:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

