

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022119\
 Data File : VX007640.D
 Acq On : 21 Feb 2019 19:22
 Operator : JC/SP
 Sample : PB117250ZHE#14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117250ZHE#14

Quant Time: Feb 22 06:19:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	440811	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	730281	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	707403	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	331131	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	255757	54.66	ug/l	0.00
Spiked Amount			Recovery	=	109.32%	
35) Dibromofluoromethane	5.51	113	241590	50.84	ug/l	0.00
Spiked Amount			Recovery	=	101.68%	
50) Toluene-d8	8.71	98	893885	50.70	ug/l	0.00
Spiked Amount			Recovery	=	101.40%	
62) 4-Bromofluorobenzene	11.13	95	309921	48.15	ug/l	0.00
Spiked Amount			Recovery	=	96.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	26565	12.395	ug/l	94
18) Methyl Acetate	2.78	43	13464	2.976	ug/l	95
20) Methylene Chloride	2.86	84	12561	2.496	ug/l	97
25) 2-Butanone	4.70	43	5356	1.756	ug/l #	88
43) Isopropyl Acetate	6.46	43	24176	2.365	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022119\
Data File : VX007640.D
Acq On : 21 Feb 2019 19:22
Operator : JC/SP
Sample : PB117250ZHE#14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117250ZHE#14

Quant Time: Feb 22 06:19:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

