

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022123\
 Data File : VX034199.D
 Acq On : 21 Feb 2023 09:57
 Operator : JC/MD
 Sample : VX0221WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0221WBL01

Quant Time: Feb 21 15:52:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	324849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	537831	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	478626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	215666	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	189184	43.243	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.480%
35) Dibromofluoromethane	5.391	113	168740	48.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	8.647	98	645937	51.376	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.760%
62) 4-Bromofluorobenzene	11.079	95	223489	46.472	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022123\
Data File : VX034199.D
Acq On : 21 Feb 2023 09:57
Operator : JC/MD
Sample : VX0221WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0221WBL01

Quant Time: Feb 21 15:52:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

