

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022123\
 Data File : VX034200.D
 Acq On : 21 Feb 2023 10:33
 Operator : JC/MD
 Sample : VX0221WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0221WBL02

Quant Time: Feb 21 15:52:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	324896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	538224	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	479180	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	218700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	189677	43.349	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.700%
35) Dibromofluoromethane	5.385	113	169834	48.464	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.920%
50) Toluene-d8	8.647	98	647413	51.456	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.920%
62) 4-Bromofluorobenzene	11.079	95	227049	47.178	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022123\
Data File : VX034200.D
Acq On : 21 Feb 2023 10:33
Operator : JC/MD
Sample : VX0221WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0221WBL02

Quant Time: Feb 21 15:52:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

