

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022123\
 Data File : VX034205.D
 Acq On : 21 Feb 2023 12:45
 Operator : JC/MD
 Sample : 01611-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2023-2-17

Quant Time: Feb 22 05:49:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	308109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	509854	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	452007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	195465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	180622	43.529	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.060%
35) Dibromofluoromethane	5.385	113	159244	47.970	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.940%
50) Toluene-d8	8.647	98	605581	50.809	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.079	95	204340	44.822	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.640%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	18143	8.476	ug/l	89
20) Methylene Chloride	2.794	84	4222	1.095	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022123\
Data File : VX034205.D
Acq On : 21 Feb 2023 12:45
Operator : JC/MD
Sample : 01611-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2023-2-17

Quant Time: Feb 22 05:49:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

