

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX022124\  
 Data File : VX040372.D  
 Acq On : 21 Feb 2024 11:36  
 Operator : JC/MD  
 Sample : VX0221WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0221WBL01

Quant Time: Feb 22 01:04:34 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020924W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Feb 12 00:21:14 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 80699    | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 143497   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 135368   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.018 | 152   | 69573    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 62394    | 51.971   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 103.940% |          |
| 35) Dibromofluoromethane    | 5.385  | 113   | 47011    | 46.912   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 93.820%  |          |
| 50) Toluene-d8              | 8.647  | 98    | 176215   | 47.009   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 94.020%  |          |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 68098    | 44.543   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 89.080%  |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX022124\  
Data File : VX040372.D  
Acq On : 21 Feb 2024 11:36  
Operator : JC/MD  
Sample : VX0221WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0221WBL01

Quant Time: Feb 22 01:04:34 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020924W.M  
Quant Title : SW846 8260  
QLast Update : Mon Feb 12 00:21:14 2024  
Response via : Initial Calibration

