

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
 Data File : VX040376.D
 Acq On : 21 Feb 2024 13:11
 Operator : JC/MD
 Sample : P1520-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP04-20240215

Quant Time: Feb 22 01:06:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	145214	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63289	52.393	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.780%
35) Dibromofluoromethane	5.385	113	47670	47.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.020%
50) Toluene-d8	8.647	98	178613	47.085	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.160%
62) 4-Bromofluorobenzene	11.079	95	68494	44.272	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.540%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	519	0.603	ug/l #	81
16) Acetone	2.373	43	1488	2.999	ug/l	91
44) Trichloroethene	7.129	130	29977	27.356	ug/l	97
64) Tetrachloroethene	9.275	164	2091	2.164	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
Data File : VX040376.D
Acq On : 21 Feb 2024 13:11
Operator : JC/MD
Sample : P1520-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP04-20240215

Quant Time: Feb 22 01:06:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

