

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
 Data File : VX040378.D
 Acq On : 21 Feb 2024 13:56
 Operator : JC/MD
 Sample : P1521-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CG-MW-23

Quant Time: Feb 22 01:07:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	77117	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	137400	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60296	52.556	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.120%	
35) Dibromofluoromethane	5.385	113	45849	47.783	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.560%	
50) Toluene-d8	8.647	98	171164	47.687	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.380%	
62) 4-Bromofluorobenzene	11.079	95	65137	44.497	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
Data File : VX040378.D
Acq On : 21 Feb 2024 13:56
Operator : JC/MD
Sample : P1521-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CG-MW-23

Quant Time: Feb 22 01:07:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

