

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
 Data File : VX040383.D
 Acq On : 21 Feb 2024 15:51
 Operator : JC/MD
 Sample : P1520-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178S-20240215

Quant Time: Feb 22 01:10:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131071	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65212	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58935	54.267	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.540%
35) Dibromofluoromethane	5.385	113	43255	47.256	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.520%
50) Toluene-d8	8.647	98	159451	46.569	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.140%
62) 4-Bromofluorobenzene	11.079	95	64605	46.264	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.520%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	911	2.042	ug/l #	86
19) Methyl tert-butyl Ether	3.111	73	1734	0.660	ug/l #	91
30) Chloroform	5.093	83	632	0.383	ug/l #	83
44) Trichloroethene	7.123	130	428	0.433	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
Data File : VX040383.D
Acq On : 21 Feb 2024 15:51
Operator : JC/MD
Sample : P1520-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178S-20240215

Quant Time: Feb 22 01:10:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

