

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
 Data File : VX040385.D
 Acq On : 21 Feb 2024 16:36
 Operator : JC/MD
 Sample : P1520-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D2-20240216

Quant Time: Feb 22 01:10:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71863	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130745	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	62713	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57747	54.014	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 108.020%		
35) Dibromofluoromethane	5.385	113	43287	47.409	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 94.820%		
50) Toluene-d8	8.647	98	159058	46.570	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 93.140%		
62) 4-Bromofluorobenzene	11.079	95	62524	44.886	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 89.780%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	7529	9.880	ug/l	98
12) 1,1-Dichloroethene	2.319	96	1699	2.316	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	4436	4.654	ug/l	92
30) Chloroform	5.093	83	699	0.430	ug/l	99
44) Trichloroethene	7.123	130	249752	253.140	ug/l	93
64) Tetrachloroethene	9.275	164	8483	9.615	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
Data File : VX040385.D
Acq On : 21 Feb 2024 16:36
Operator : JC/MD
Sample : P1520-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D2-20240216

Quant Time: Feb 22 01:10:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

