

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
 Data File : VX040386.D
 Acq On : 21 Feb 2024 16:59
 Operator : JC/MD
 Sample : P1532-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D2-20240216

Quant Time: Feb 22 01:11:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70460	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	125973	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	62521	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56059	53.479	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.960%
35) Dibromofluoromethane	5.385	113	41588	47.274	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.540%
50) Toluene-d8	8.646	98	156946	47.692	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.380%
62) 4-Bromofluorobenzene	11.079	95	61657	45.940	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.880%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	2911	3.896	ug/l	97
12) 1,1-Dichloroethene	2.318	96	309	0.430	ug/l #	73
16) Acetone	2.385	43	932	2.165	ug/l #	81
44) Trichloroethene	7.122	130	295839	311.211	ug/l	97
64) Tetrachloroethene	9.274	164	871	1.023	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
Data File : VX040386.D
Acq On : 21 Feb 2024 16:59
Operator : JC/MD
Sample : P1532-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D2-20240216

Quant Time: Feb 22 01:11:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

