

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
 Data File : VX040387.D
 Acq On : 21 Feb 2024 17:22
 Operator : JC/MD
 Sample : P1532-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20240216

Quant Time: Feb 22 01:11:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	69413	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	124163	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60768	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	55645	53.885	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.780%
35) Dibromofluoromethane	5.385	113	41518	47.882	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.760%
50) Toluene-d8	8.647	98	153648	47.371	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.740%
62) 4-Bromofluorobenzene	11.079	95	59522	44.996	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.000%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	1284	1.744	ug/l #	88
16) Acetone	2.380	43	921	2.172	ug/l	99
44) Trichloroethene	7.123	130	125490	133.935	ug/l	100
64) Tetrachloroethene	9.275	164	428	0.508	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
Data File : VX040387.D
Acq On : 21 Feb 2024 17:22
Operator : JC/MD
Sample : P1532-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20240216

Quant Time: Feb 22 01:11:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

