

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
 Data File : VX040390.D
 Acq On : 21 Feb 2024 18:31
 Operator : JC/MD
 Sample : P1532-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D3-20240216

Quant Time: Feb 22 01:12:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123504	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55351	55.732	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.460%
35) Dibromofluoromethane	5.385	113	41991	48.686	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.380%
50) Toluene-d8	8.647	98	150210	46.558	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.120%
62) 4-Bromofluorobenzene	11.079	95	56912	43.252	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.500%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	1722	4.222	ug/l	95
44) Trichloroethene	7.129	130	2241	2.405	ug/l	99
52) Toluene	8.714	92	2579	1.165	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
Data File : VX040390.D
Acq On : 21 Feb 2024 18:31
Operator : JC/MD
Sample : P1532-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D3-20240216

Quant Time: Feb 22 01:12:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

