



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022124\
Data File : VX040369.D
Acq On : 21 Feb 2024 08:19
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Title : SW846 8260
Last Update : Mon Feb 12 00:21:14 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.7	5770	PASS
75	95	30	60	52.9	14078	PASS
95	95	100	100	100.0	26616	PASS
96	95	5	9	6.5	1726	PASS
173	174	0.00	2	1.1	215	PASS
174	95	50	100	72.6	19333	PASS
175	174	5	9	6.7	1301	PASS
176	174	95	101	95.7	18502	PASS
177	176	5	9	6.5	1197	PASS