

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022219\
 Data File : VX007659.D
 Acq On : 22 Feb 2019 14:56
 Operator : JC/SP
 Sample : K1570-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 10-10

Quant Time: Feb 23 01:14:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X021319W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Feb 13 07:51:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	66766	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	331463	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	292146	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	133399	30.89	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.97%
60) 4-Bromofluorobenzene	11.13	95	123826	27.62	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.07%
63) Toluene-d8	8.71	98	392804	29.75	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.17%

Target Compounds					Ovalue
15) Acetone	2.45	58	7412	11.718	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022219\
Data File : VX007659.D
Acq On : 22 Feb 2019 14:56
Operator : JC/SP
Sample : K1570-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
10-10

Quant Time: Feb 23 01:14:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X021319W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Feb 13 07:51:18 2019
Response via : Initial Calibration

