


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022219\
Data File : VX007645.D
Acq On : 22 Feb 2019 08:00
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Title : SW846 8260
Last Update : Sat Feb 02 01:32:36 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	6249	PASS
75	95	30	60	49.2	18336	PASS
95	95	100	100	100.0	37285	PASS
96	95	5	9	7.0	2620	PASS
173	174	0.00	2	1.5	493	PASS
174	95	50	100	88.6	33034	PASS
175	174	5	9	7.4	2456	PASS
176	174	95	101	97.3	32133	PASS
177	176	5	9	6.6	2126	PASS