

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022222\
 Data File : VX027127.D
 Acq On : 22 Feb 2022 11:10
 Operator : JC/MD
 Sample : VX0222MBL01
 Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0222MBL01

Quant Time: Feb 22 15:15:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	106389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173381	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67531	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62940	49.284	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.560%
35) Dibromofluoromethane	5.391	113	55155	53.395	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.780%
50) Toluene-d8	8.647	98	195877	51.659	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.320%
62) 4-Bromofluorobenzene	11.079	95	70728	52.982	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022222\
Data File : VX027127.D
Acq On : 22 Feb 2022 11:10
Operator : JC/MD
Sample : VX0222MBL01
Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0222MBL01

Quant Time: Feb 22 15:15:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

