

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022222\
 Data File : VX027128.D
 Acq On : 22 Feb 2022 11:33
 Operator : JC/MD
 Sample : VX0222WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0222WBL01

Quant Time: Feb 22 15:16:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	100444	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	168133	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150865	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60914	50.520	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.040%	
35) Dibromofluoromethane	5.391	113	53066	52.976	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.960%	
50) Toluene-d8	8.653	98	190786	51.887	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.780%	
62) 4-Bromofluorobenzene	11.079	95	71483	55.219	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.440%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022222\
Data File : VX027128.D
Acq On : 22 Feb 2022 11:33
Operator : JC/MD
Sample : VX0222WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0222WBL01

Quant Time: Feb 22 15:16:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

