

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022223\
 Data File : VX034220.D
 Acq On : 22 Feb 2023 11:26
 Operator : JC/MD
 Sample : VX0222WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0222WBL01

Quant Time: Feb 23 01:26:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	304117	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	501675	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	440818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	190386	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	178299	43.533	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.060%
35) Dibromofluoromethane	5.385	113	157580	48.243	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.480%
50) Toluene-d8	8.647	98	594850	50.723	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.440%
62) 4-Bromofluorobenzene	11.079	95	197831	44.101	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022223\
Data File : VX034220.D
Acq On : 22 Feb 2023 11:26
Operator : JC/MD
Sample : VX0222WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0222WBL01

Quant Time: Feb 23 01:26:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

