

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
 Data File : VX040397.D
 Acq On : 22 Feb 2024 13:27
 Operator : JC/MD
 Sample : VX0222WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0222WBL01

Quant Time: Feb 23 00:05:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76259	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	134163	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60143	53.012	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.020%	
35) Dibromofluoromethane	5.385	113	44700	47.709	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.420%	
50) Toluene-d8	8.647	98	164901	47.051	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 94.100%	
62) 4-Bromofluorobenzene	11.079	95	63861	44.678	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
Data File : VX040397.D
Acq On : 22 Feb 2024 13:27
Operator : JC/MD
Sample : VX0222WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0222WBL01

Quant Time: Feb 23 00:05:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

