

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
 Data File : VX040400.D
 Acq On : 22 Feb 2024 14:40
 Operator : JC/MD
 Sample : P1520-09DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D1-20240216DL

Quant Time: Feb 23 00:07:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72600	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	125274	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56054	51.898	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.800%
35) Dibromofluoromethane	5.385	113	42007	48.016	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.040%
50) Toluene-d8	8.647	98	154938	47.345	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.680%
62) 4-Bromofluorobenzene	11.079	95	58249	43.643	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.280%
Target Compounds						
44) Trichloroethene	7.129	130	32135	33.993	ug/l	99
64) Tetrachloroethene	9.275	164	1824	2.177	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
Data File : VX040400.D
Acq On : 22 Feb 2024 14:40
Operator : JC/MD
Sample : P1520-09DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D1-20240216DL

Quant Time: Feb 23 00:07:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

