

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
 Data File : VX040401.D
 Acq On : 22 Feb 2024 15:03
 Operator : JC/MD
 Sample : P1520-10DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20240216DL

Quant Time: Feb 23 00:08:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	127701	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122500	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57047	52.912	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.820%
35) Dibromofluoromethane	5.385	113	42920	48.128	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.260%
50) Toluene-d8	8.647	98	157968	47.353	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.700%
62) 4-Bromofluorobenzene	11.079	95	60710	44.622	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.240%
Target Compounds						
						Qvalue
44) Trichloroethene	7.123	130	29144	30.243	ug/l	93
64) Tetrachloroethene	9.281	164	1696	1.976	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
Data File : VX040401.D
Acq On : 22 Feb 2024 15:03
Operator : JC/MD
Sample : P1520-10DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20240216DL

Quant Time: Feb 23 00:08:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

