

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
 Data File : VX040404.D
 Acq On : 22 Feb 2024 16:12
 Operator : JC/MD
 Sample : P1532-02 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20240216

Quant Time: Feb 23 00:09:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	68736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	125118	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56215	54.973	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.940%	
35) Dibromofluoromethane	5.385	113	41126	47.068	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.140%	
50) Toluene-d8	8.647	98	154823	47.369	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 94.740%	
62) 4-Bromofluorobenzene	11.079	95	59322	44.502	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.000%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	662	0.908	ug/l #	89
44) Trichloroethene	7.129	130	48360	51.220	ug/l	98
64) Tetrachloroethene	9.281	164	614	0.715	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
Data File : VX040404.D
Acq On : 22 Feb 2024 16:12
Operator : JC/MD
Sample : P1532-02 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20240216

Quant Time: Feb 23 00:09:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

