

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
 Data File : VX040405.D
 Acq On : 22 Feb 2024 16:35
 Operator : JC/MD
 Sample : P1532-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D2-20240216DL

Quant Time: Feb 23 00:09:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	69260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123178	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58879	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55584	53.945	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.880%			
35) Dibromofluoromethane	5.385	113	40762	47.386	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.780%			
50) Toluene-d8	8.647	98	150849	46.880	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 93.760%			
62) 4-Bromofluorobenzene	11.079	95	58668	44.705	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.400%			
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.501	96	4614	5.022	ug/l	94
44) Trichloroethene	7.123	130	39860	42.883	ug/l	97
52) Toluene	8.726	92	859	0.389	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
Data File : VX040405.D
Acq On : 22 Feb 2024 16:35
Operator : JC/MD
Sample : P1532-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D2-20240216DL

Quant Time: Feb 23 00:09:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

