

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
 Data File : VX040406.D
 Acq On : 22 Feb 2024 16:58
 Operator : JC/MD
 Sample : P1532-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D3-20240216

Quant Time: Feb 23 00:09:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	131379	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60153	55.475	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.940%
35) Dibromofluoromethane	5.385	113	43981	47.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.880%
50) Toluene-d8	8.647	98	160955	46.898	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.800%
62) 4-Bromofluorobenzene	11.079	95	61936	44.249	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.500%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1673	3.757	ug/l #	84
44) Trichloroethene	7.129	130	895	0.903	ug/l	86
52) Toluene	8.720	92	2604	1.106	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
Data File : VX040406.D
Acq On : 22 Feb 2024 16:58
Operator : JC/MD
Sample : P1532-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D3-20240216

Quant Time: Feb 23 00:09:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

