

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
 Data File : VX040417.D
 Acq On : 22 Feb 2024 23:51
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 23 06:34:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 23 06:27:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74555	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	129479	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59507	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57954	47.731	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.460%
35) Dibromofluoromethane	5.385	113	43426	49.770	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.540%
50) Toluene-d8	8.647	98	158265	50.216	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.440%
62) 4-Bromofluorobenzene	11.079	95	61707	48.861	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
Data File : VX040417.D
Acq On : 22 Feb 2024 23:51
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 23 06:34:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 23 06:27:42 2024
Response via : Initial Calibration

