



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022224\
Data File : VX040409.D
Acq On : 22 Feb 2024 18:30
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
Title : SW846 8260
Last Update : Fri Feb 23 06:47:29 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.4	6041	PASS
75	95	30	60	56.7	14653	PASS
95	95	100	100	100.0	25845	PASS
96	95	5	9	6.8	1767	PASS
173	174	0.00	2	0.6	106	PASS
174	95	50	100	73.0	18865	PASS
175	174	5	9	7.1	1340	PASS
176	174	95	101	95.5	18010	PASS
177	176	5	9	6.0	1088	PASS