

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
 Data File : VX027143.D
 Acq On : 23 Feb 2022 14:01
 Operator : JC/MD
 Sample : VX0223MBL01
 Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0223MBL01

Quant Time: Feb 24 03:30:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	101162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	166195	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68660	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59587	49.044	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.080%
35) Dibromofluoromethane	5.385	113	51440	49.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.720%
50) Toluene-d8	8.653	98	190786	49.461	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.079	95	71609	48.364	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
Data File : VX027143.D
Acq On : 23 Feb 2022 14:01
Operator : JC/MD
Sample : VX0223MBL01
Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0223MBL01

Quant Time: Feb 24 03:30:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

