

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
 Data File : VX027149.D
 Acq On : 23 Feb 2022 16:35
 Operator : JC/MD
 Sample : N1616-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 24 03:32:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	92855	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	153377	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63177	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55115	49.422	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.840%
35) Dibromofluoromethane	5.391	113	47717	50.121	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.240%
50) Toluene-d8	8.653	98	178307	50.089	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.180%
62) 4-Bromofluorobenzene	11.079	95	66797	48.884	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
Data File : VX027149.D
Acq On : 23 Feb 2022 16:35
Operator : JC/MD
Sample : N1616-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 24 03:32:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

