

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
 Data File : VX027152.D
 Acq On : 23 Feb 2022 17:46
 Operator : JC/MD
 Sample : N1648-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-1

Quant Time: Feb 24 03:32:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	90640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	151104	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54121	49.716	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.440%
35) Dibromofluoromethane	5.397	113	46484	49.560	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.120%
50) Toluene-d8	8.652	98	173855	49.573	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.140%
62) 4-Bromofluorobenzene	11.085	95	66002	49.029	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.060%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	9321	19.502	ug/l	94
25) 2-Butanone	4.580	43	2793	3.669	ug/l	93
51) 4-Methyl-2-Pentanone	8.579	43	5707	3.798	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
Data File : VX027152.D
Acq On : 23 Feb 2022 17:46
Operator : JC/MD
Sample : N1648-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-1

Quant Time: Feb 24 03:32:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

