

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
 Data File : VX027154.D
 Acq On : 23 Feb 2022 18:33
 Operator : JC/MD
 Sample : N1648-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-3

Quant Time: Feb 24 03:33:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	91075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	149756	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64423	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53360	48.783	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.560%
35) Dibromofluoromethane	5.391	113	46424	49.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.880%
50) Toluene-d8	8.653	98	172478	49.623	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.240%
62) 4-Bromofluorobenzene	11.079	95	65013	48.729	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	11465	23.873	ug/l	91
25) 2-Butanone	4.574	43	3260	4.262	ug/l	100
51) 4-Methyl-2-Pentanone	8.580	43	11127	7.471	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
Data File : VX027154.D
Acq On : 23 Feb 2022 18:33
Operator : JC/MD
Sample : N1648-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-3

Quant Time: Feb 24 03:33:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

