

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
 Data File : VX027155.D
 Acq On : 23 Feb 2022 18:56
 Operator : JC/MD
 Sample : N1647-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31382

Quant Time: Feb 24 03:33:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	89331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	147595	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52728	49.147	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.300%
35) Dibromofluoromethane	5.391	113	45639	49.816	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.640%
50) Toluene-d8	8.653	98	168788	49.273	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.540%
62) 4-Bromofluorobenzene	11.079	95	63079	47.972	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.940%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1159	2.460	ug/l	96
52) Toluene	8.720	92	2599	1.038	ug/l	100
68) m/p-Xylenes	10.305	106	1456	0.801	ug/l	96
69) o-Xylene	10.647	106	747	0.410	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	1303	0.356	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	5426	1.489	ug/l	100
95) Naphthalene	13.780	128	1458	0.341	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
Data File : VX027155.D
Acq On : 23 Feb 2022 18:56
Operator : JC/MD
Sample : N1647-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31382

Quant Time: Feb 24 03:33:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

