

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
 Data File : VX027142.D
 Acq On : 23 Feb 2022 13:26
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX022322

Quant Time: Feb 24 03:30:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.497	0.505	-1.6	105	0.00
3 P	Chloromethane	0.465	0.464	0.2	100	0.00
4 C	Vinyl Chloride	0.463	0.474	-2.4#	102	0.00
5 T	Bromomethane	0.210	0.223	-6.2	122	0.00
6 T	Chloroethane	0.284	0.289	-1.8	104	0.00
7 T	Trichlorofluoromethane	0.720	0.750	-4.2	105	0.00
8 T	Diethyl Ether	0.251	0.268	-6.8	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.501	0.536	-7.0	107	0.00
10 T	Methyl Iodide	0.673	0.688	-2.2	99	0.00
11 T	Tert butyl alcohol	0.117	0.122	-4.3	106	0.00
12 CM	1,1-Dichloroethene	0.487	0.515	-5.7#	106	0.00
13 T	Acrolein	0.108	0.108	0.0	104	0.00
14 T	Allyl chloride	0.857	0.917	-7.0	106	0.00
15 T	Acrylonitrile	0.300	0.318	-6.0	104	0.00
16 T	Acetone	0.264	0.305	-15.5	121	0.00
17 T	Carbon Disulfide	1.266	1.325	-4.7	105	0.00
18 T	Methyl Acetate	0.681	0.724	-6.3	105	0.00
19 T	Methyl tert-butyl Ether	1.670	1.768	-5.9	104	0.00
20 T	Methylene Chloride	0.546	0.559	-2.4	104	0.00
21 T	trans-1,2-Dichloroethene	0.519	0.549	-5.8	105	0.00
22 T	Diisopropyl ether	1.671	1.774	-6.2	104	0.00
23 T	Vinyl Acetate	1.411	1.552	-10.0	105	0.00
24 P	1,1-Dichloroethane	0.944	0.988	-4.7	104	0.00
25 T	2-Butanone	0.420	0.457	-8.8	108	0.00
26 T	2,2-Dichloropropane	0.685	0.782	-14.2	112	0.00
27 T	cis-1,2-Dichloroethene	0.602	0.646	-7.3	106	0.00
28 T	Bromochloromethane	0.382	0.367	3.9	94	0.00
29 T	Tetrahydrofuran	0.281	0.291	-3.6	103	0.00
30 C	Chloroform	0.981	1.012	-3.2#	103	0.00
31 T	Cyclohexane	0.899	0.935	-4.0	104	0.00
32 T	1,1,1-Trichloroethane	0.849	0.903	-6.4	104	0.00
33 S	1,2-Dichloroethane-d4	0.601	0.565	6.0	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.310	0.295	4.8	97	0.00
36 T	1,1-Dichloropropene	0.453	0.484	-6.8	104	0.00
37 T	Ethyl Acetate	0.497	0.539	-8.5	107	0.00
38 T	Carbon Tetrachloride	0.472	0.514	-8.9	104	0.00
39 T	Methylcyclohexane	0.563	0.622	-10.5	108	0.00
40 TM	Benzene	1.325	1.413	-6.6	104	0.00
41 T	Methacrylonitrile	0.273	0.299	-9.5	106	0.00
42 TM	1,2-Dichloroethane	0.470	0.506	-7.7	105	0.00
43 T	Isopropyl Acetate	0.765	0.849	-11.0	106	0.00
44 TM	Trichloroethene	0.396	0.423	-6.8	104	0.00
45 C	1,2-Dichloropropane	0.324	0.355	-9.6#	104	0.00
46 T	Dibromomethane	0.232	0.252	-8.6	104	0.00
47 T	Bromodichloromethane	0.454	0.486	-7.0	104	0.00
48 T	Methyl methacrylate	0.370	0.409	-10.5	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
 Data File : VX027142.D
 Acq On : 23 Feb 2022 13:26
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX022322

Quant Time: Feb 24 03:30:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	104	0.00
50 S	Toluene-d8	1.160	1.105	4.7	97	0.00
51 T	4-Methyl-2-Pentanone	0.497	0.529	-6.4	102	0.00
52 CM	Toluene	0.848	0.902	-6.4#	103	0.00
53 T	t-1,3-Dichloropropene	0.446	0.514	-15.2	105	0.00
54 T	cis-1,3-Dichloropropene	0.506	0.574	-13.4	105	0.00
55 T	1,1,2-Trichloroethane	0.336	0.354	-5.4	101	0.00
56 T	Ethyl methacrylate	0.523	0.570	-9.0	102	0.00
57 T	1,3-Dichloropropane	0.559	0.597	-6.8	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.252	0.290	-15.1	109	0.00
59 T	2-Hexanone	0.380	0.407	-7.1	102	0.00
60 T	Dibromochloromethane	0.348	0.383	-10.1	103	0.00
61 T	1,2-Dibromoethane	0.358	0.382	-6.7	101	0.00
62 S	4-Bromofluorobenzene	0.445	0.410	7.9	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.452	0.490	-8.4	103	0.00
65 PM	Chlorobenzene	1.019	1.081	-6.1	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.402	-8.6	103	0.00
67 C	Ethyl Benzene	1.789	1.941	-8.5#	103	0.00
68 T	m/p-Xylenes	0.688	0.749	-8.9	101	0.00
69 T	o-Xylene	0.690	0.733	-6.2	101	0.00
70 T	Styrene	1.130	1.237	-9.5	101	0.00
71 P	Bromoform	0.280	0.307	-9.6	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.578	3.878	-8.4	102	0.00
74 T	N-amyl acetate	1.361	1.546	-13.6	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.057	1.108	-4.8	102	0.00
76 T	1,2,3-Trichloropropane	1.007	1.077	-7.0	101	0.00
77 T	Bromobenzene	0.892	0.947	-6.2	100	0.00
78 T	n-propylbenzene	4.066	4.415	-8.6	101	0.00
79 T	2-Chlorotoluene	2.492	2.642	-6.0	101	0.00
80 T	1,3,5-Trimethylbenzene	3.008	3.239	-7.7	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.315	0.349	-10.8	104	0.00
82 T	4-Chlorotoluene	2.847	3.026	-6.3	100	0.00
83 T	tert-Butylbenzene	2.925	3.125	-6.8	100	0.00
84 T	1,2,4-Trimethylbenzene	2.993	3.265	-9.1	102	0.00
85 T	sec-Butylbenzene	3.668	4.023	-9.7	102	0.00
86 T	p-Isopropyltoluene	3.127	3.458	-10.6	102	0.00
87 T	1,3-Dichlorobenzene	1.679	1.777	-5.8	100	0.00
88 T	1,4-Dichlorobenzene	1.717	1.794	-4.5	101	0.00
89 T	n-Butylbenzene	2.614	2.905	-11.1	102	0.00
90 T	Hexachloroethane	0.484	0.542	-12.0	102	0.00
91 T	1,2-Dichlorobenzene	1.638	1.733	-5.8	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.245	0.272	-11.0	102	0.00
93 T	1,2,4-Trichlorobenzene	1.034	1.144	-10.6	103	0.00
94 T	Hexachlorobutadiene	0.429	0.471	-9.8	103	0.00
95 T	Naphthalene	3.513	3.891	-10.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022322\
Data File : VX027142.D
Acq On : 23 Feb 2022 13:26
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX022322

Quant Time: Feb 24 03:30:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.036	1.111	-7.2	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6