

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022323\
 Data File : VX034245.D
 Acq On : 23 Feb 2023 13:26
 Operator : JC/MD
 Sample : 01666-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10-EB-20230221

Quant Time: Feb 23 21:42:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	282217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	464510	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	407666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178149	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163819	43.101	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.200%
35) Dibromofluoromethane	5.385	113	146163	48.328	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.660%
50) Toluene-d8	8.647	98	545662	50.251	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.500%
62) 4-Bromofluorobenzene	11.079	95	184829	44.500	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.000%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	14317	16.842	ug/l #	81
16) Acetone	2.380	43	21618	11.026	ug/l #	88
52) Toluene	8.720	92	2102	0.262	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022323\
Data File : VX034245.D
Acq On : 23 Feb 2023 13:26
Operator : JC/MD
Sample : 01666-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-EB-20230221

Quant Time: Feb 23 21:42:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

