

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022323\
 Data File : VX034248.D
 Acq On : 23 Feb 2023 15:02
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 23 21:44:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	276558	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	456587	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	404168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	182352	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	164081	44.054	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.100%
35) Dibromofluoromethane	5.391	113	142319	47.873	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.740%
50) Toluene-d8	8.647	98	543620	50.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.860%
62) 4-Bromofluorobenzene	11.079	95	185072	45.331	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022323\
Data File : VX034248.D
Acq On : 23 Feb 2023 15:02
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 23 21:44:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

