

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022323\
 Data File : VX034250.D
 Acq On : 23 Feb 2023 15:51
 Operator : JC/MD
 Sample : 01650-01
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BU-2-022223

Quant Time: Feb 23 21:46:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	249978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	412951	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367923	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	183049	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	143920	42.749	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.500%
35) Dibromofluoromethane	5.385	113	124955	46.474	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.940%
50) Toluene-d8	8.647	98	488809	50.636	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.280%
62) 4-Bromofluorobenzene	11.085	95	177649	48.111	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.220%
Target Compounds						
11) Tert butyl alcohol	2.947	59	6545	8.692	ug/l	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022323\
Data File : VX034250.D
Acq On : 23 Feb 2023 15:51
Operator : JC/MD
Sample : 01650-01
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BU-2-022223

Quant Time: Feb 23 21:46:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

