

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022323\
 Data File : VX034244.D
 Acq On : 23 Feb 2023 13:03
 Operator : JC/MD
 Sample : 01666-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10-TB-20230220

Quant Time: Feb 23 21:42:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	291709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	476511	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	413047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	177151	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	169073	43.036	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.080%
35) Dibromofluoromethane	5.385	113	149341	48.135	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.280%
50) Toluene-d8	8.647	98	567820	50.975	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.940%
62) 4-Bromofluorobenzene	11.079	95	188959	44.348	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.700%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022323\
Data File : VX034244.D
Acq On : 23 Feb 2023 13:03
Operator : JC/MD
Sample : 01666-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-TB-20230220

Quant Time: Feb 23 21:42:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX034244.D
Acq: 23 Feb 2023 13:03




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-TB-20230220


Tgt Ion:168 Resp: 291709
Ion Ratio Lower Upper
168 100
99 53.1 39.6 59.4



#33
1,2-Dichloroethane-d4
Concen: 43.036 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX034244.D
Acq: 23 Feb 2023 13:03

Tgt Ion: 65 Resp: 169073
Ion Ratio Lower Upper
65 100
67 53.2 0.0 104.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. 0.006 min
Lab File: VX034244.D
Acq: 23 Feb 2023 13:03




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-TB-20230220


Tgt Ion:114 Resp: 476511
Ion Ratio Lower Upper
114 100
63 19.4 0.0 37.4
88 15.8 0.0 30.6



#35
Dibromofluoromethane
Concen: 48.135 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX034244.D
Acq: 23 Feb 2023 13:03

Tgt Ion:113 Resp: 149341
Ion Ratio Lower Upper
113 100
111 102.9 83.0 124.6
192 20.4 19.0 28.4









#50
Toluene-d8
Concen: 50.975 ug/l
RT: 8.647 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034244.D
Acq: 23 Feb 2023 13:03




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-TB-20230220


Tgt Ion: 98 Resp: 567820
Ion Ratio Lower Upper
98 100
100 66.0 53.6 80.4



#62
4-Bromofluorobenzene
Concen: 44.348 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX034244.D
Acq: 23 Feb 2023 13:03

Tgt Ion: 95 Resp: 188959
Ion Ratio Lower Upper
95 100
174 81.4 0.0 184.4
176 78.5 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX034244.D
Acq: 23 Feb 2023 13:03

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-TB-20230220

Tgt Ion:117 Resp: 413047
Ion Ratio Lower Upper
117 100
82 53.2 40.6 60.8
119 32.4 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX034244.D
Acq: 23 Feb 2023 13:03

Tgt Ion:152 Resp: 177151
Ion Ratio Lower Upper
152 100
115 57.9 38.7 116.1
150 155.0 0.0 348.2

