


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022323\
Data File : VX034237.D
Acq On : 23 Feb 2023 08:48
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Title : SW846 8260
Last Update : Fri Feb 17 12:33:35 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	10388	PASS
75	95	30	60	49.7	26997	PASS
95	95	100	100	100.0	54317	PASS
96	95	5	9	6.6	3563	PASS
173	174	0.00	2	0.8	338	PASS
174	95	50	100	79.2	43037	PASS
175	174	5	9	8.2	3549	PASS
176	174	95	101	95.6	41136	PASS
177	176	5	9	7.1	2913	PASS